



DIVISION OF INDUSTRY SERVICES
Plumbing Product Review
P.O. Box 2658
Madison, Wisconsin 53701-2658
TTY: Contact Through Relay

Scott Walker, Governor
Dave Ross, Secretary

October 27, 2014

MEDISAFE AMERICA LLC
GARY JORDAN
9423 CORPORATE LAKE DRIVE
TAMPA FL 33634

MEDISAFE UK LIMITED
UMAR NAGI
BISHOPS STORTFORD, HERTFORDSHIRE
CM23 3LJ UNITED KINGDOM

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: MEDISAFE UK LIMITED
Product Name: MEDISAFE SIPCF WASHER DISINFECTOR
Model Number(s): MED11121
Product File No: 20140245

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes. The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of October 2019. This approval is contingent upon compliance with the following stipulation(s):

- A reduced pressure principle backflow preventer (i.e. RPZ) shall be installed in the hot and cold water supply lines serving these devices. The RPZ shall conform to American Society of Sanitary Engineers (ASSE) Standard 1013.
- The RPZ's shall be installed and maintained as specified in s. SPS 382.41 (4) and s. SPS 382.41 (5) Wis. Adm. Code.
- The RPZ's shall be registered with this department as specified in s. SPS 382.20 (1) (c) and s. SPS 382.20 (13) Wis. Adm. Code.
- The RPZ's shall be tested as specified in s. SPS 382.22 (8) Wis. Adm. Code.
- The drain, waste and vent (i.e. DWV) piping serving these devices must be properly sized to accept the maximum rate of discharge from these devices. The preexisting DWV system must be sufficiently sized to accept the additional wastewater loading generated by these devices.
- The DWV piping serving these devices must be comprised of a material(s) that are stable at the wastewater temperatures generated by these devices.
- These devices shall discharge wastewater in accordance with s. SPS 382.50 (2) (a) 2.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter
Environmental Engineer - Plumbing Product Reviewer
Department of Safety and Professional Services
Division of Industry Services
Bureau of Technical Services
(608) 267-1401 Phone
(608) 267-9723 Fax
glen.schlueter@wi.gov E-mail